


Analysis of Linear alkylbenzene sulfonate

Data No. LB151-0888



Conditions

System : LC800 HPLC System
4000 QTRAP®

Column : InertSustain C18 HP
(3 µm, 150×2.1 mm I.D.)

Column Cat. No. : 5020-14415

Eluent : A) CH₃CN
B) 0.1 % HCOOH + 50 mM HCOONH₄ in H₂O
A/B = 65/35, v/v

Flow Rate : 0.2 mL/min.

Col. Temp. : 40 °C

Detection : LC/MS/MS (4000 QTRAP® ESI, Negative, MRM)

CUR	CAD	IS	TEM	GS1	GS2
10	4	-4500	600	70	40

Injection Vol. : 5 µL

Analyte:

1. Sodium Octylbenzenesulfonate(C8) (I.S.)
 2. Sodium Decylbenzenesulfonate(C10)
 3. Sodium Undecylbenzenesulfonate(C11)
 4. Sodium Dodecylbenzenesulfonate(C12)
 5. Sodium Tridecylbenzenesulfonate(C13)
 6. Sodium Tetradecylbenzenesulfonate(C14)
- in CH₃CN/H₂O = 65/35, v/v
each 5 µg/L