


InertSearch™ for LC

Inertsil® Applications

Analysis of Divinylbenzene

Data No. LB170-0521



Conditions

System : GL-7400 HPLC system
Column : Inertsil C30 S-Select
(5 μ m, 250 x 4.6 mm I.D.)
Column Cat. No. : 5020-17037
Eluent : A) CH₃CN
B) H₂O
A/B = 60/40, v/v
Flow Rate : 1.0 mL/min
Col. Temp. : 40 °C
Detection : UV 254 nm (GL-7452 PDA Detector)
Injection Vol. : 1 μ L
Sample : Standard

Analyte:

1. *m*-, *p*-Divinylbenzene
2. *m*-, *p*-Ethylvinylbenzene