


Analysis of Fexofenadine hydrochloride

(Under the Condition of the Japanese Pharmacopoeia)

Data No. LB134-0919



Conditions

System : GL-7400 HPLC system

Column : Inertsil Ph
(5 μ m, 250 x 4.6 mm I.D.)

Column Cat. No. : 5020-01328

Eluent : A) CH₃CN
B) 7.51 g/L NaH₂PO₄ • 2H₂O + 0.84 g/L NaClO₄ in H₂O (pH 2.0, H₃PO₄)
C) Triethylamine
A/B/C = 350/650/3, v/v/v

Flow Rate : 1.8 mL/min

Col. Temp. : 25 °C

Detection : UV 220 nm (GL-7452 PDA Detector)

Injection Vol. : 20 μ L

Sample : Standard

Analyte:

1. Fexofenadine hydrochloride