


Analysis of Melamine related compounds

Data No. LB500-806

**Conditions**

Column : InertSustain Amide
(5 μ m, 150 x 2.1 mm I.D.)
Column Cat. No. : 5020-88607
Eluent : A) CH₃CN
B) 10 mM HCOONH₄ in H₂O
A/B = 90/10, v/v
Flow Rate : 0.2 mL/min
Col. Temp. : 40 °C
Detection : UV 215 nm
Injection Vol. : 0.5 μ L
Sample : Standard

Analyte:

1. Cyanuric acid	1 mg/mL
2. Melamine	0.1 mg/mL
3. Ammelide	0.1 mg/mL
4. Ammeline	0.1 mg/mL