

D-Mannitol Analysis Under USP Method


United States Pharmacopoeia (USP), European Pharmacopoeia (EP), and Japanese Pharmacopoeia (JP) have harmonized the analytical method for Mannitol. As a result of the harmonization, a HPLC method was added.

In this note, chromatograms obtained with an InertSphere Sugar-2 column, which was developed for sugar analysis, according to the harmonized method are shown. Whereas the original Japanese version of this note was written according to the JP, in this English version, the results are shown with reference to the USP monograph.


(Y. Yui, R. Hirano)

Analyte: 1,3. Isomalt
2. Maltitol
4. D-Mannitol
5. D-Sorbitol

System suitability solution A



System suitability solution B



HPLC conditions

Column : InertSphere Sugar-2 (9 mm, 300 x 7.8 mm I.D.)
Eluent : H₂O
Col. Temp. : 85 °C
Detection : RI (40 °C)
Injection Vol. : 20 µL
Flow rate : 0.5 mL/min

System suitability (Assay and Related Substances)





Resolution (4, 5) : 5.46 (> 2.0)

Relative retention time (Related Substances)

Isomalt (1st peak) : 0.66 (0.60)*
Maltitol : 0.71 (0.69)*
Isomalt (2nd peak) : 0.74 (0.73)*
Sorbitol : 1.19 (1.2)*

* Reference value written in USP monograph

Analyte: 4. Mannitol

Standard solution AStandard solution BStandard solution B
(Enlarged chromatogram)Standard solution C

** Negative peak detected at about 18 min is increased or decreased depending on dissolved air in eluent and sample solution.

Analytical Column for Mannitol analysis

USP Code: L19

Description: Strong cation-exchange resin consisting of sulfonated cross-linked styrene-divinylbenzene copolymer in the calcium form, about 9 µm in diameter.

InertSphere Sugar-2

Description : InertSphere Sugar-2 (9 µm, 300 x 7.8 mm I.D.)
 Cat.No. : 5020-11000
 Base Material : Styrene-divinylbenzene copolymer
 Functional Group : Sulfonic acid
 Counter Ion : Ca²⁺



GL Sciences disclaims any and all responsibility for any injury or damage which may be caused by this data directly or indirectly. We reserve the right to amend this information or data at any time and without any prior announcement.

GL Sciences Inc. Japan

22-1 Nishishinjuku 6-chome
 Shinjuku-ku, Tokyo
 163-1130, Japan

Phone: +81-3-5323-6620
 Fax: +81-3-5323-6621
 Email: world@glsc.co.jp
 Web: www.glsciences.com

GL Sciences Inc. USA

4733 Torrance Blvd. Suite 255
 Torrance, CA 90503
 USA

Phone: +1-310-265-4424
 Fax: +1-310-265-4425
 Email: info@glsciencesinc.com
 Web: www.glsciencesinc.com

GL Sciences B.V.

Dillenburgstraat 7C
 5652AM, Eindhoven
 The Netherlands

Phone: +31-40-254-9531
 Email: info@glsciences.eu
 Web: www.glsciences.eu

GL Sciences (Shanghai) Limited

Tower B, Room 2003
 Far East International Plaza
 No.317 Xianxia Road, Changning District
 Shanghai, China 200051

Phone: +86-21-62782272
 Email: contact@glsciences.com.cn
 Web: www.glsciences.com.cn



International Distributors

Visit our Website at www.glsciences.com/distributors